

Refurbish Report Sheet

Classification		Item				Result	Spec	Judge
8	B/C Scope Confirmation	1	TOC Measurement Repeatability	Mode1	B	0.011	$\leq$ 3 σ 0.020 μ m	ok
					C	0.010		
				Mode2	B			
					C			
				Mode3	B	0.013		
					C	0.013		
				Mode4	B			
					C			
		2	Defocus Characteristics	B(Mode1)	Ka	-1	$\pm$ 20	ok
					Kb	-2		
					Kc	0		
				C(Mode1)	Ka	2	$\pm$ 20	
					Kb	4		
					Kc	0		
				B(Mode2)	Ka	-14	$\pm$ 20	
					Kb	-12		
					Kc	0		
				C(Mode2)	Ka	-3	$\pm$ 20	
					Kb	-3		
					Kc	0		
				B(Mode3)	Ka	0	$\pm$ 20	
					Kb	0		
					Kc	0		
				C(Mode3)	Ka	5	$\pm$ 20	
					Kb	6		
					Kc	0		
				B(Mode4)	Ka	-1	$\pm$ 20	
					Kb	-1		
					Kc	0		
				C(Mode4)	Ka	2	$\pm$ 20	
					Kb	3		
					Kc	0		