

B96AF20-B

Final Inspection Report

C-PTD 150/XL/AIR/18P

Identification			
166901	C-PTD 150/XL/AIR/18P	Serial number	
	CCB - Connector board	Serial number	

Standard delivery items				
Part number	Pcs.	Part name	o.k.	Acronym
166901	1	C-PTD 150/XL/AIR/18P	☒	1295

Alignment check				
Checkpoint	Max. deviation	Equipment	Meas. value	Acronym
Parallelism	$\pm 10 \mu\text{m}$	10668	$\pm 7 \mu\text{m}$	1295
Concentricity	$\pm 5 \mu\text{m}$	12122	$\pm 5 \mu\text{m}$	1295

Temperature check				
Checkpoint	Max. deviation	Equipment	Meas. value	Acronym
+20 °C	$\pm 0.2 \text{ °C}$	7891	20.01 °C	1295

Temperature profile			
Equipment	Set temperature	o.k.	Acronym
8872	+20 °C	☒	1295
	+10 °C	☒	1295
	+150 °C	☒	1295

Additional checks		
Checkpoint	Equipment	Acronym
Electrical check	-	1295
Fan function	-	1295
Visual inspection	-	1295

Device Identification		
Programming	o.k.	Acronym
Toolmaster™	☒	1295

Date	Checked by	Signature
28.03.2019		

C-PTD 150/XL/AIR/18P



Peltier Performance C-PTD 150/XL/AIR/18P; T: 20,01°C; SN: 82815707

— T

Anton Paar